REDACTED COA COPY

Supplier identifiers removed; analytical data unchanged.

In-house report; not an independent third-party test.

CERTIFICATE NUMBER

PL-26-00057

LOT / BATCH

PL-2609-057

CERTIFICATE OF ANALYSIS · LOT RELEASE

KLOW Blend 80 mg

Lot **PL-2609-057** · Lyophilized co-blend powder · Label claim **80 mg/vial**

IN-HOUSE ANALYTICAL REPORT

In-house analytical result, not an independent third-party test

LOT / BATCH NUMBER	DATE MADE	DATE OF ANALYSIS	DATE ISSUED
PL-2609-057	2026-09-15	2026-09-16	2026-10-03
RETEST / EXPIRY	**UNITS IN LOT / SAMPLED**	**STORAGE**	**CERTIFICATE TYPE**
2028-09-15	250 / 2	−20 °C, desiccated, dark	Lot COA, 2 pages

COMPOUND

KLOW Blend
GHK-Cu + BPC-157 + TB-500 + KPV co-lyophilized blend

CAS NUMBER

89030-95-5 · 137525-51-0 · 77591-33-4 · 67727-97-3

MOLECULAR FORMULA

$C_{14}H_{22}CuN_6O_4$ · $C_{62}H_{98}N_{16}O_{22}$ · $C_{212}H_{350}N_{56}O_{78}S$ · $C_{16}H_{30}N_4O_4$

THEORETICAL MASS (AVG.)

401.9 · 1419.6 · 4963.5 · 342.4 Da (GHK-Cu / BPC-157 / TB-500 (Thymosin β4) / KPV)

COMPOSITION AND SEQUENCES

GHK-Cu 50 mg: `H-GHK-OH · Cu(II) (1:1 copper complex)` · **BPC-157 10 mg:** `H-GEPPPGKPADDAGLV-OH (15 residues; free N- and C-termini)` · **TB-500 (Thymosin β4) 10 mg:** `Ac-SDKPDMAEIEKFDKSKLKKTETQEKNPLPSKETIEQEKQAGES-OH (43 residues; N-acetyl; free C-terminus)` · **KPV 10 mg:** `H-KPV-OH (3 residues; free N- and C-termini)`

PURITY · RP-HPLC, AREA %

99.66%

Specification 99.0% • blend purity — sum of the four component peaks

IDENTITY · OBSERVED MASS

4 / 4

Components confirmed by LC-MS within ± 1.0 Da: 402.1 / 1419.8 / 4963.7 / 342.6 Da

OVERALL CONCLUSION


PASS

Meets every specification for the tests performed

TEST	METHOD	SPECIFICATION	RESULT	CONCLUSION
Appearance	Visual inspection	Blue to blue-violet lyophilized cake (GHK-Cu copper complex)	Blue lyophilized cake	**PASS**
Identity (mass)	LC-MS, ESI positive	Theoretical ± 1.0 Da, each component	4/4 components confirmed	**PASS**
Purity	RP-HPLC, UV 214 nm, area%	99.0% (blend)	99.66%	**PASS**
Largest impurity	RP-HPLC, area%	1.0%	0.11% at 5.35 min	**PASS**
Water content	Karl Fischer	5.0%	2.1%	**PASS**
Bacterial endotoxin	LAL, kinetic	≤ 5.0 EU/mg	0.06 EU/mg	**PASS**
Sterility / bioburden	USP <71>, membrane filtration	No growth after 14 days	No growth; conforms	**PASS**

1 HPLC CHROMATOGRAM

Data file: KLOW_PL-2609-057_20260916_01 · Column C18 150 × 4.6 mm · UV 214 nm



Baseline uncropped; four component peaks and impurities labeled with retention times. Peak table on page 2.

For laboratory research use only. Not for human or veterinary use.

2 HPLC METHOD · KLOW_PL-2609-057_20260916_01

INSTRUMENT
Waters Alliance e2695, SN E2695B742

DATA SYSTEM
Empower 3, Build 3471

DETECTOR
Waters 2998 PDA

DETECTION WAVELENGTH
214 nm

COLUMN
C18, 150 × 4.6 mm, 3.5 µm

COLUMN PART / SERIAL
186003326 / USHX18211

MOBILE PHASE A
0.1% TFA in water

MOBILE PHASE B
0.1% TFA in MeCN

FLOW RATE
1.0 mL/min

COLUMN TEMPERATURE
30 °C

INJECTION VOLUME
20 µL

RUN TIME
10 min

SAMPLE DILUENT
Water, 0.1% TFA

SAMPLE CONCENTRATION
0.5 mg/mL (blend)

BLANK INJECTED
Yes

REFERENCE STANDARD
In-house KLOW blend reference, lot PL-RS-KLO-01, component RTs 2.18 / 3.29 / 4.66 / 6.06 min

3 GRADIENT PROGRAM

TIME (MIN)	% A	% B
0.0	95	5
3.0	95	5
7.0	55	45
8.0	20	80
9.0	20	80
10.0	95	5

4 PEAK TABLE

PEAK	RT (MIN)	AREA	AREA %
1 (GHK-Cu)	2.20	2,682,340	62.38
2	2.85	3,870	0.09
3 (KPV)	3.31	525,890	12.23
4 (BPC-157)	4.68	531,910	12.37
5	5.35	4,730	0.11
6 (TB-500)	6.08	545,240	12.68
7	6.90	3,440	0.08
8	7.62	2,580	0.06
Total		4,300,000	100.00

Threshold 0.05% area · Blank subtracted: Y · Purity = main peak area ÷ total integrated area × 100, to two decimals; main peak shaded.

5 IDENTITY BY MASS SPECTROMETRY

INSTRUMENT
Waters Acquity SQD

IONIZATION / MODE
ESI, positive

OBSERVED M/Z (CHARGE)
402.8 · 1420.3 (1+ [M+H], GHK-Cu · BPC-157) · 343.2 (1+ [M+H], KPV) · 1241.9 (4+ [M+4H], TB-500) · 993.7 (5+ [M+5H], TB-500)

DECONVOLUTED MASS
402.1 / 1419.8 / 4963.7 / 342.6 Da (by component)

THEORETICAL MASS
401.9 / 1419.6 / 4963.5 / 342.4 Da (GHK-Cu / BPC-157 / TB-500 / KPV)

DIFFERENCE
≤ +0.2 Da (per component)

DATA FILE
KLOW_PL-2609-057_MS_20260916_01

ACCEPTANCE
Within ± 1.0 Da (each component)

6 ADDITIONAL TESTS

TEST	PERFORMED BY	DATE	REPORT / FILE
Water content	In-house	2026-09-17	RPT-KF-PL-2609-057
Endotoxin	In-house	2026-09-18	RPT-ET-PL-2609-057
Sterility	In-house	2026-09-30	RPT-ST-PL-2609-057

A test with no row here was not performed on this lot. Outside-laboratory reports, when used, are attached and named below.

7 REVIEW AND APPROVAL

D. Reynolds
Analyst · performed the testing
Date: 2026-09-16

S. Calloway
Reviewer · checked against raw data
Date: 2026-10-02

M. Phelps
Lab Director · approved for release
Date: 2026-10-03

Raw data retained at: Empower 3 (Build 3471), project PL-2609, archive PL-ARC-2026; 7 years. Outside laboratory reports attached: None

Redacted copy: supplier name, logo, address, contact details and verification QR/URL have been removed. This certificate applies only to the named lot and sampled units. Results were obtained by the supplier's in-house laboratory and are not independent third-party testing unless an outside report is attached. Reference values (CAS, formula, theoretical mass, sequence) are literature values. For laboratory research use only; not for human or veterinary use.