REDACTED COA COPY

Supplier identifiers removed; analytical data unchanged.

In-house report; not an independent third-party test.

CERTIFICATE NUMBER
PL-26-00064

LOT / BATCH
PL-2609-064

CERTIFICATE OF ANALYSIS · LOT RELEASE

MOTS-c 40 mg

Lot **PL-2609-064** · Lyophilized powder · Label claim **40 mg/vial**

IN-HOUSE ANALYTICAL REPORT

In-house analytical result, not an independent third-party test

LOT / BATCH NUMBER	DATE MADE	DATE OF ANALYSIS	DATE ISSUED
PL-2609-064	2026-09-05	2026-09-07	2026-10-03
RETEST / EXPIRY	**UNITS IN LOT / SAMPLED**	**STORAGE**	**CERTIFICATE TYPE**
2028-09-05	250 / 2	−20 °C, desiccated, dark	Lot COA, 2 pages

COMPOUND
MOTS-c
Mitochondrial open reading frame of the 12S rRNA type-c

CAS NUMBER
1627580-64-6

MOLECULAR FORMULA
$C_{101}H_{152}N_{28}O_{22}S_2$

THEORETICAL MASS (AVG.)
2174.6 Da

SEQUENCE
`H-MRWQEMGYIFYPRKLR-OH (16 residues; free N- and C-termini)`

PURITY · RP-HPLC, AREA %
99.56%
Specification 99.0% • main peak at 6.44 min


IDENTITY · OBSERVED MASS
2174.8 Da
Theoretical 2174.6 Da • +0.2 Da (92 ppm)

OVERALL CONCLUSION
PASS
Meets every specification for the tests performed

TEST	METHOD	SPECIFICATION	RESULT	CONCLUSION
Appearance	Visual inspection	White to off-white lyophilized powder	White lyophilized powder	PASS
Identity (mass)	LC-MS, ESI positive	Theoretical ± 1.0 Da	2174.8 Da	PASS
Purity	RP-HPLC, UV 214 nm, area%	99.0%	99.56%	PASS
Largest impurity	RP-HPLC, area%	1.0%	0.17% at 7.22 min	PASS
Water content	Karl Fischer	5.0%	1.7%	PASS
Bacterial endotoxin	LAL, kinetic	≤ 5.0 EU/mg	0.06 EU/mg	PASS
Sterility / bioburden	USP <71>, membrane filtration	No growth after 14 days	No growth; conforms	PASS

1 HPLC CHROMATOGRAM

Data file: MOTS_PL-2609-064_20260907_01 · Column C18 150 × 4.6 mm · UV 214 nm



Baseline uncropped; main peak and impurities labeled with retention times. Peak table on page 2.

For laboratory research use only. Not for human or veterinary use.

2 HPLC METHOD · MOTS_PL-2609-064_20260907_01

INSTRUMENT
Waters Alliance e2695, SN E2695B742

DATA SYSTEM
Empower 3, Build 3471

DETECTOR
Waters 2998 PDA

DETECTION WAVELENGTH
214 nm

COLUMN
C18, 150 × 4.6 mm, 3.5 µm

COLUMN PART / SERIAL
186003326 / USHX18211

MOBILE PHASE A
0.1% TFA in water

MOBILE PHASE B
0.1% TFA in MeCN

FLOW RATE
1.0 mL/min

COLUMN TEMPERATURE
30 °C

INJECTION VOLUME
20 µL

RUN TIME
10 min

SAMPLE DILUENT
Water, 0.1% TFA

SAMPLE CONCENTRATION
0.5 mg/mL

BLANK INJECTED
Yes

REFERENCE STANDARD
In-house MOTS-c reference, lot PL-RS-MOT-01, RT 6.42 min

3 GRADIENT PROGRAM

TIME (MIN)	% A	% B
0.0	95	5
3.0	95	5
7.0	55	45
8.0	20	80
9.0	20	80
10.0	95	5

4 PEAK TABLE

PEAK	RT (MIN)	AREA	AREA %
1	5.61	6,525	0.15
2	**6.44**	**4,330,860**	**99.56**
3	7.22	7,395	0.17
4	7.98	5,220	0.12
Total		4,350,000	100.00

Threshold 0.05% area · Blank subtracted: Y · Purity = main peak area ÷ total integrated area × 100, to two decimals; main peak shaded.

5 IDENTITY BY MASS SPECTROMETRY

INSTRUMENT
Waters Acquity SQD

IONIZATION / MODE
ESI, positive

OBSERVED M/Z (CHARGE)
1088.4 (2+ [M+2H]) · 725.9 (3+ [M+3H])

DECONVOLUTED MASS
2174.8 Da

THEORETICAL MASS
2174.6 Da

DIFFERENCE
+0.2 Da (92 ppm)

DATA FILE
MOTS_PL-2609-064_MS_20260907_01

ACCEPTANCE
Within ± 1.0 Da

6 ADDITIONAL TESTS

TEST	PERFORMED BY	DATE	REPORT / FILE
Water content	In-house	2026-09-08	RPT-KF-PL-2609-064
Endotoxin	In-house	2026-09-09	RPT-ET-PL-2609-064
Sterility	In-house	2026-09-21	RPT-ST-PL-2609-064

A test with no row here was not performed on this lot. Outside-laboratory reports, when used, are attached and named below.

7 REVIEW AND APPROVAL

D. Reynolds
Analyst · performed the testing
Date: 2026-09-07

S. Calloway
Reviewer · checked against raw data
Date: 2026-10-02

M. Phelps
Lab Director · approved for release
Date: 2026-10-03

Raw data retained at: Empower 3 (Build 3471), project PL-2609, archive PL-ARC-2026; 7 years. Outside laboratory reports attached: None

Redacted copy: supplier name, logo, address, contact details and verification QR/URL have been removed. This certificate applies only to the named lot and sampled units. Results were obtained by the supplier's in-house laboratory and are not independent third-party testing unless an outside report is attached. Reference values (CAS, formula, theoretical mass, sequence) are literature values. For laboratory research use only; not for human or veterinary use.