REDACTED COA COPY

Supplier identifiers removed; analytical data unchanged.

In-house report; not an independent third-party test.

CERTIFICATE NUMBER

PL-26-00098

LOT / BATCH

PL-2609-098

CERTIFICATE OF ANALYSIS · LOT RELEASE

Sermorelin 10 mg

Lot **PL-2609-098** · Lyophilized powder · Label claim **10 mg/vial**

IN-HOUSE ANALYTICAL REPORT

In-house analytical result, not an independent third-party test

LOT / BATCH NUMBER	DATE MADE	DATE OF ANALYSIS	DATE ISSUED
PL-2609-098	2026-09-18	2026-09-19	2026-10-03
RETEST / EXPIRY	**UNITS IN LOT / SAMPLED**	**STORAGE**	**CERTIFICATE TYPE**
2028-09-18	250 / 2	−20 °C, desiccated, dark	Lot COA, 2 pages

COMPOUND	CAS NUMBER	MOLECULAR FORMULA	THEORETICAL MASS (AVG.)
Sermorelin GHRH(1–29)-NH_2; GRF(1–29)	86168-78-7	$C_{149}H_{246}N_{44}O_{42}S$	3357.9 Da

SEQUENCE

`H-YADAIFTNSYRKVLGQLSARKLLQDIMSR-NH2 (29 residues; C-terminal amide)`

PURITY · RP-HPLC, AREA %

99.63%

Specification 99.0% • main peak at 6.34 min

IDENTITY · OBSERVED MASS

3358.1 Da

Theoretical 3357.9 Da • +0.2 Da (60 ppm)

OVERALL CONCLUSION


PASS

Meets every specification for the tests performed

TEST	METHOD	SPECIFICATION	RESULT	CONCLUSION
Appearance	Visual inspection	White to off-white lyophilized powder	White lyophilized powder	PASS
Identity (mass)	LC-MS, ESI positive	Theoretical ± 1.0 Da	3358.1 Da	PASS
Purity	RP-HPLC, UV 214 nm, area%	99.0%	99.63%	PASS
Largest impurity	RP-HPLC, area%	1.0%	0.19% at 7.12 min (des-Arg 29 fragment)	PASS
Water content	Karl Fischer	5.0%	1.8%	PASS
Bacterial endotoxin	LAL, kinetic	≤ 5.0 EU/mg	0.05 EU/mg	PASS
Sterility / bioburden	USP <71>, membrane filtration	No growth after 14 days	No growth; conforms	PASS

1 HPLC CHROMATOGRAM

Data file: SRM_PL-2609-098_20260919_01 · Column C18 150 × 4.6 mm · UV 214 nm



Baseline uncropped; main peak and impurities labeled with retention times. Peak table on page 2.

For laboratory research use only. Not for human or veterinary use.

2 HPLC METHOD · SRM_PL-2609-098_20260919_01

INSTRUMENT
Waters Alliance e2695, SN E2695B742

DATA SYSTEM
Empower 3, Build 3471

DETECTOR
Waters 2998 PDA

DETECTION WAVELENGTH
214 nm

COLUMN
C18, 150 × 4.6 mm, 3.5 µm

COLUMN PART / SERIAL
186003326 / USHX18211

MOBILE PHASE A
0.1% TFA in water

MOBILE PHASE B
0.1% TFA in MeCN

FLOW RATE
1.0 mL/min

COLUMN TEMPERATURE
30 °C

INJECTION VOLUME
20 µL

RUN TIME
10 min

SAMPLE DILUENT
Water, 0.1% TFA

SAMPLE CONCENTRATION
0.5 mg/mL

BLANK INJECTED
Yes

REFERENCE STANDARD
In-house Sermorelin reference, lot PL-RS-SRM-01, RT 6.32 min

3 GRADIENT PROGRAM

TIME (MIN)	% A	% B
0.0	95	5
3.0	95	5
7.0	55	45
8.0	20	80
9.0	20	80
10.0	95	5

4 PEAK TABLE

PEAK	RT (MIN)	AREA	AREA %
1	5.52	6,020	0.14
2	**6.34**	**4,284,090**	**99.63**
3	6.82	6,450	0.15
4 (des-Arg)	7.12	8,170	0.19
5	7.88	1,720	0.04
Total		4,306,450	100.00 (0.15 rounding)

Threshold 0.05% area · Blank subtracted: Y · Purity = main peak area ÷ total integrated area × 100, to two decimals; main peak shaded.

5 IDENTITY BY MASS SPECTROMETRY

INSTRUMENT
Waters Acquity SQD

IONIZATION / MODE
ESI, positive

OBSERVED M/Z (CHARGE)
1120.4 (3+ [M+3H]) · 840.5 (4+ [M+4H])

DECONVOLUTED MASS
3358.1 Da

THEORETICAL MASS
3357.9 Da

DIFFERENCE
+0.2 Da (60 ppm)

DATA FILE
SRM_PL-2609-098_MS_20260919_01

ACCEPTANCE
Within ± 1.0 Da

6 ADDITIONAL TESTS

TEST	PERFORMED BY	DATE	REPORT / FILE
Water content	In-house	2026-09-20	RPT-KF-PL-2609-098
Endotoxin	In-house	2026-09-21	RPT-ET-PL-2609-098
Sterility	In-house	2026-10-02	RPT-ST-PL-2609-098

A test with no row here was not performed on this lot. Outside-laboratory reports, when used, are attached and named below.

7 REVIEW AND APPROVAL

D. Reynolds
Analyst · performed the testing
Date: 2026-09-19

S. Calloway
Reviewer · checked against raw data
Date: 2026-10-02

M. Phelps
Lab Director · approved for release
Date: 2026-10-03

Raw data retained at: Empower 3 (Build 3471), project PL-2609, archive PL-ARC-2026; 7 years. Outside laboratory reports attached: None

Redacted copy: supplier name, logo, address, contact details and verification QR/URL have been removed. This certificate applies only to the named lot and sampled units. Results were obtained by the supplier's in-house laboratory and are not independent third-party testing unless an outside report is attached. Reference values (CAS, formula, theoretical mass, sequence) are literature values. For laboratory research use only; not for human or veterinary use.